

O2k-Manual: TIP2k Manual



Mitochondrial Physiology Network 12.10(09):1-20 (2021)

Version 09: 2021-07-08 ©2007-2021 Oroboros

Updates: http://wiki.orooboros.at/index.php/MiPNet12.10_TIP2k-Manual

Titration-Injection microPump TIP2k manual

Lukas Gradl³, Philip Gradl², Erich Gnaiger¹

¹Oroboros Instruments


High-Resolution Respirometry
Schoepfstrasse 18, 6020 Innsbruck, Austria
Email: instruments@orooboros.at
www.orooboros.at

²WGT Elektronik GmbH & Co KG

Rettenbergstraße 30a
6114 Kolsass, Austria

³ssn software security networks

Eduard-Bodem-Gasse 5
6020 Innsbruck, Austria



Contents

1.	TIP2k main features – TIP2k-Module	2
2.	Set up the TIP2k.....	2
2.1.	TIP2k components	2
2.2.	Installation.....	3
2.3.	Manual operation	3
2.4.	Safety.....	3
3.	Filling and mounting the syringes	4
3.1.	Filling the TIP2k syringe	4
3.2.	Mounting the TIP2k syringe.....	5
3.3.	Inserting the TIP2k needle into the O2k-chamber:	6
3.4.	Washing the TIP2k needle after use	6
3.5.	Preventing TIP2k needle clogging	7
4.	Operation of the TIP2k with DatLab.....	7
4.1.	TIP2k control [F8]: quick start and introduction	7
4.2.	TIP2k control window - overview	8
4.3.	Structure of a TIP2k setup	8
4.4.	Creating and editing TIP2k setups	9
4.5.	Parameters for a program line.....	9
4.6.	Direct control mode:	11
4.7.	Feedback control mode.....	11
4.8.	Save, load and manage TIP2k setups.....	15
4.9.	Running a TIP2k setup program	16

4.10. Chemicals window.....	17
4.11. TIP2k configuration window.....	17
4.12. Info window.....	17
4.13. TIP2k status.....	18
4.14. TIP2k events.....	19
5. References.....	20
6. Author contributions.....	20
7. Acknowledgements.....	20

1. TIP2k main features – TIP2k-Module

2-channel Titration-Injection microPump: The TIP2k provides unique advantages for applications which require highest accuracy of titration volumes and automatic control of complex titration-injection protocols. The TIP2k was specially developed for the Oroboros O2k. The main modes of operation are automatic multi-step injections, continuous injections for steady-state respirometry, (both operated by the direct control mode), and feedback controlled operation.

Voltage supply: The power supply of the TIP2k and electronic control is integrated into the Oroboros O2k.

Direct control mode: Operation mode for delivering a defined volume in a short pulse. The number of cycles, N, can be defined, and the duration of the interval from injection to injection determined, repeating the process N times. Several such steps can be programmed to run in sequence to give an entire TIP2k program.

Feedback control mode: Operation mode to reach and maintain a certain parameter (e.g. oxygen concentration) by automatic TIP injections.

2. Set up the TIP2k

Operation of the TIP2k is exclusively possible in conjunction with the O2k and DatLab.

2.1. TIP2k components

Dimensions: D x W x H: 270 x 124 x 160 mm (without syringes); stainless steel housing, 1.5 mm – 2.0 mm thick. Weight: 2.7 kg.

TIP2k microsyringes: Simultaneous operation of two 200 μL (mm^3) Hamilton syringes; 0.12 mm inner diameter flexible stainless-steel needles, 180 mm length, with TIP2k-Needle Spacers. For larger volumes, 500 μL microsyringes are available.

All programmed volumes and flows will be calculated erroneously if the syringe mounted does not match the specifications in the DatLab “TIP Control” window.

TIP2k-Needle Spacer: Package of silicone stops (200/Pkg.) with mounting tool.



2.2. Installation

Placement: The TIP2k is placed on the top of the O2k. It does not need any extra bench space. When using the 200 μ L microsyringes with the TIP2k, make sure to place the TIP2k correctly. The rubber feet of the TIP2k should be positioned 5 cm from the front edge of the O2k (see picture). If you received the TIP2k separately, use the tool for placing the TIP2k-Position marker correctly on the O2k (included in the shipment). Of note, when using the 500 μ L microsyringes, the TIP2k must be positioned more than 5 cm from the front edge of the O2k.

Environment: The TIP2k was tested at room temperature and must be protected from temperature fluctuations and condensation of water. The instrument must not be exposed to chemicals or vibrations, and the syringes should not be exposed to direct sunlight during operation.

Cable: A 0.5 m cable connects the TIP2k to the O2k at the rear, with a 15 pin female SubD plug.

When the TIP2k is connected to the O2k and the main unit of the Oroboros O2k is switched on, the blue (rear push button) and green (front push button) lights on the side of the TIP2k light up, indicating proper connection (TIP2k C and higher). In TIP2k A and B series there is a single light on the rear push button on the side of the TIP2k.



2.3. Manual operation

Syringe holders (blue POM) are mounted on each side of the TIP, for a syringe with its needle guided into the left and right O2k-chamber. The front part of the holders is static, for fixation of the glass cylinder of the 200- μ L microsyringe, while the rear part is moved with the step motor of the TIP2k, for fixation of the plunger. When using 200- μ L microsyringes with the TIP2k, tighten the back screw and do not use the front screw. Otherwise, the syringe may break. It is important to have the screw of the syringe plunger pointing towards the side to avoid damaging the syringe and the TIP2k. Of note, when using 500- μ L microsyringe tighten also the front screw in TIP2k series A or B.



Push buttons are located on the right side of the TIP2k. Push the front one (green light) for forwards movement, and the rear one (blue light) for backwards movement of the plunger holder. For stopping the manual movement, simply release the push button. Pushing both buttons simultaneously also stops the pump. At manual operation of the TIP2k, the step motor moves at maximum speed.

Front and rear limit: When the front or rear limit of the pump is reached, the motor stops automatically. This status is indicated by the blinking blue and green light on the rear and front push button, respectively. After a short push on the TIP2k button in the other direction, the TIP2k is in the status ready, the blue and the green light are on without blinking.

2.4. Safety

Potential hazards:



When not inserted into the O2k-chamber, the needles of the TIP2k syringes may - due to their length and flexibility - inflict serious injuries to the eyes.

- Dangerous chemicals may be splashed out of the syringes by unintentionally operating the TIP2k.
- If a blocked syringe is used in the TIP2k, the syringe may fracture under the applied pressure and serious injury may occur, especially to the eyes by glass fragments.

Safety rules:

- **Always wear safety goggles!**
- **Never operate the TIP2k with a blocked syringe!**
- When operating the TIP2k without the needles inserted into the O2k-chamber: Place the end of the needle securely into the tube of the TIP2k-Needle Safety Support to avoid any hazards to persons or instruments by jetting the solution out of the needle!



3. Filling and mounting the syringes

Dust and any remains in the syringe cause clogging of the needle as well as wearing of the plunger. When using a blocked needle, the pump may be damaged due to the generation of high pressure. Therefore, carefully read and observe the instructions for washing the needles and prevention of needle clogging.



Any inclusion of gas bubbles in the syringe must be avoided. Solutions must not be subjected to an increase in temperature during or after filling. Otherwise, gassing out will give rise to bubbles which cause erroneous titration volumes or undesirable hysteresis effects.



Do not over-bend the stainless steel needles, since this causes irreversible damage.

3.1. Filling the TIP2k syringe

Before filling the syringe, make sure that it is clean and rinsed with the solvent in which the substance to be titrated-injected is dissolved. Filling the TIP2k syringe can be done by back-filling after removal of the plunger, or through the needle while pulling the plunger backwards. The backfilling of the TIP2k syringe is faster when enough solution is available for completely filling the syringe. If injection solution is limiting, it is possible to fill the syringe with a small volume (e.g. 50 μL) using suction force.

The different ways to fill the syringes are:

Filling the TIP2k syringe from the glass cylinder – requires $>200 \mu\text{L}$ solution per microsyringe: Remove the plunger from the syringe. Fix the syringe in vertical position on a holder. Using a standard injection syringe with a long needle, fill the TIP2k syringe from bottom to top, without inclusion of gas bubbles at the bottom or on the glass wall. The plunger is inserted while a small amount of the solution is first extruded on the top. Once a tight fitting of the plunger is obtained, the solution is extruded through the needle. The small diameter and length of the needle cause a substantial resistance against ejection of the solution. Place the end of the needle securely into a beaker, to avoid any hazards to persons or instruments by jetting the solution out of the needle.

Filling the TIP2k syringe through the needle – is possible with $<200 \mu\text{L}$ solution:

1. Push the plunger completely into the TIP2k syringe.
2. Immerse the tip of the needle into the titration solution and pull the plunger slowly back to the desired volume mark. It is advisable to suck in some extra


- volume since some solution may be spilled while attempting to remove any trapped air bubbles.
3. Allow for a slow pressure equilibration, until most of the gas space is filled by the solution. A small gas volume fills the space between the plunger and the solution.
 4. Hold the syringe vertically with the needle pointed up and gently tap at the side of the syringe at the point where the air bubble is trapped. The bubble moves up the syringe to the needle outlet. Position the bubble at the center of the glass syringe, directly at the opening to the needle, by gently tapping at the side of the syringe.
 5. Slightly press the plunger to squeeze out the bubble, whereby some solution may be lost from the syringe. If the bubble moves to the side while pressing the plunger, stop pushing and gently tap the syringe tip until the bubble is located centrally, and continue. Avoid any undesired contact with the titration solution, and prevent the long needle from moving uncontrolled, which may be dangerous for the eyes.
 6. After the bubble has been displaced from the glass part of the syringe, eject further about 10 - 20 μL to flush the entire needle with solution.

Filling the syringe with the TIP2k

1. Before mounting the syringe, make sure that it is clean and rinsed with the solvent in which the substance to be titrated-injected is dissolved.
2. Push the plunger fully into the syringe.
3. Move the pump to the front limit. Fix the plunger heads and glass cylinders of syringes with the screws of the syringe holders.
4. Insert the tips of the needles into the titration-injection solution.
5. Shortly press the rear push button of the TIP2k to change the status from "TIP: Front limit" to "TIP: Ready". The green and blue lights at the center of the TIP2k buttons stop blinking and are on.
6. In DatLab, open the TIP2k Control window [F8], and select "TIP backward".
7. Set the "Volume μL " to be filled in the syringes. It is advisable to fill some more volume than required for the titration-injection experiments. Select the flow to a low value to minimize gassing out from the solution.
8. Click on **Test start** to fill the syringe. Provide time for pressure equilibration within the syringe. For instance, if 100 μL is filled at a flow of 50 $\mu\text{L}/\text{s}$, about 60 to 90 s are required until the pressure difference is relaxed by the slow flow of solution.
9. Remove the syringes from the syringe holders and expel the trapped air bubble as described above.

3.2. Mounting the TIP2k syringe

1. Move the pump manually to or towards the rear limit.
2. Loosen the big screw on the syringe holder (green arrow) and insert the syringe. Note that the small screw on the syringe (red arrow) must face outwards.
3. Fix the plunger head with the screw of the rear part of the syringe holder, fastening it finger tight.
4. Repeat steps 2-3 with the second syringe on the other side of the TIP2k.
5. Use two fingers of the left hand to hold both syringes, while pressing the front button on the TIP2k to move the syringes forwards until the protruding edges



of the glass cylinder touch the front syringe holders. Release the push button as soon as both syringes have reached this final position.

6. Place the end of the needle securely into tube of the TIP2k-Needle Safety Support, to avoid any hazards to persons or instruments by jetting the solution out of the needle.

3.3. Inserting the TIP2k needle into the O2k-chamber:

Before inserting the needle into the O2k-chamber, rinse the surface carefully.

In DatLab, open the 'TIP2k control' window [F8], set a low volume and low flow (e.g. 1 μL and 10 $\mu\text{L/s}$), and click on **Test start** (for details see below). No events are set when using **Test start** instead of **Start** in the F8 window. Stop the pump after a small droplet of the solution appears at the tip of the needle. Remove the extruded drop, preferentially with a plastic tip, so as not to withdraw any solution from the capillary, which may occur when using a paper towel.

Place the TIP2k Filter Papers on the top of the stoppers, to avoid liquid excess on the stopper. Place the spacer of the needle, leaving a length of 55 mm to the end. This places the needle end just above the stirrer in a closed 2-mL O2k-chamber. Be careful not to touch the stirrer bar with the tip of the needle, when inserting the needle through the capillary of the PEEK stopper into the O2k-chamber. Alternatively (e.g. for working with a different chamber volume) adjust the position of the spacer interactively: Put the spacer in a short distance to the end of the needle, insert the needle into the stopper and push the needle further down to the desired depth, while visually controlling the position through the window. For the next usage, the spacer will already be in the correct position.

3.4. Washing the TIP2k needle after use

Immediately after use, the following steps should be done:

1. Remove the needles from the O2k-chamber
2. Extrude some volume from the needle (e.g. by manually pressing the forward button). Place the end of the needle securely into a beaker, to avoid any hazards to people or instruments by jetting the solution out of the needle.
3. Rinse the outside of the needle with water to remove any media. This must be done before rinsing the interior of the syringe with any non-aqueous solvents!
4. Rinse the needles several times (3-5 times) with an appropriate solvent (i.e. one in which the substances used are soluble). Usually this will be the same solvent you used during TIP2k operation.
5. Empty the syringe and store in a safe place.

It will usually not be possible to remove all solvents from the needle. Therefore, before using the needle again, remember to rinse the needle with a pure solvent. If a needle was washed with water and an ethanolic, water-insoluble substance is filled into the syringe without prior rinsing with ethanol the needle will be blocked. The same will happen in the reverse case.

If a TIP2k syringe is harder to operate than usual, start immediately with extensive cleaning procedures. Once a needle is totally blocked it is usually impossible (with exceptions) to clean it. Suitable solvents for extensive cleaning procedures depend on the substances used. Chloroform was successful for some fatty acids. An ultrasonic bath may be helpful in cleaning partially blocked syringes.

3.5. Preventing TIP2k needle clogging


Reasons for needle clogging: To prevent washing out effect when the needle is inserted into the O2k-chamber, the inner diameter of the TIP2k needle has to be very small. This means that the needle is easily blocked by small particles either precipitated from the solution or introduced with the solution. While a TIP2k syringe has to be considered as consumable, it is possible to lengthen its lifetime by following carefully the procedures suggested in this manual.

Prevention: The main way to prevent clogging of needles is to follow rigorously the cleaning procedure outlined above after each usage. Additionally, it has to be taken care that no particles are suspended in the solutions used. For not very-well soluble substances, filtering of the solution may be necessary. Small particles can also be introduced by laboratory activities, e.g. from paper towels or powdered gloves.


4. Operation of the TIP2k with DatLab

4.1. TIP2k control [F8]: quick start and introduction

Press [F8] to open the 'TIP2k control' window.



Default settings are given at the start. These settings may be edited step-by-step, or by selecting a TIP2k setup. The main TIP2k window consists of four tabs: Control, Chemicals, Configuration, and Info. They will be covered in detail in the following sections.



Some users may want to quickly run a predefined TIP2k setup program without any modifications. For these users – as a first step - the following "quick start" information may be helpful:



Load and start a TIP2k setup: Select a TIP Setup in the ▼ pull-down menu (in the left lower corner) and click on **Load setup** to load the respective control settings. The new control settings for the TIP2k are now displayed. Settings of various setups may be viewed at this stage. Start the displayed TIP2k program by clicking on **Start**.

4.2. TIP2k control window - overview

There are two principal viewing modes for the `TIP2k control` window. In the short view, only the current TIP2k setup and control buttons for starting, loading, and saving a TIP2k setup are visible.

By pressing **Show details**, an extended view is displayed that allows editing of a TIP2k setup.

By pressing **Hide details**, the view is changed to the short modus again. For the following discussion the extended view will be used.



4.3. Structure of a TIP2k setup

A TIP2k **setup** is a predefined set of instructions to be executed by the TIP2k, plus information about the chemicals used. The entire actual setup is displayed in a table (**program lines table**) in the lower part of the `TIP2k control` window. Each setup consists of at least one (or more) **program lines**.

Line	Mode	Del [s]	Vol [µl]	Flow [µl/s]	Time [s]	Duration [s]	Cycles	Feedback quantity	><	Value	DataN	Action	Pause [s]
1	F	1200	100.000	0.250	400.00	300		1A: O2 concentration [µM]	>	120.00	2	start	0
2	F	900	50.000	0.125	400.00	300		1A: O2 concentration [µM]	>	60.00	2	start	0
3	F	900	20.000	0.050	400.00	300		1A: O2 concentration [µM]	>	30.00	2	start	0
4	D	900	50.000	50.000	1.00	120	1						

Program lines are executed by the TIP2k strictly and sequentially. Each program line may lead to several injections by the TIP2k (or none at all). After the execution of a program line the TIP2k is always in stopped mode (all injections are finished) before starting the next program line.

Each program line is identified by a number in the column "**Line**". The column "**Mode**" indicates whether a particular program line uses direct control (**D**) or feedback control (**F**). The other entries into the table will be explained when discussing the different modes to program the TIP2k.

4.4. Creating and editing TIP2k setups

Editing and creating TIP2k setups must be done in the detailed view of the “Control” tab, press **Show details** if necessary. Editing of a setup can only be done when no TIP2k setup program is running, otherwise the editing functions will be locked. For editing and creating TIP2k setups a DatLab file has to be open either in on-line or off-line mode.

Editing and creating program lines is done by using the **input mask** in the upper part of the Control tab, the **control buttons** situated between the input mask and the program lines table, and the **program lines table** in the lower part of the window.

The screenshot shows the TIP2k Control window with the following parameters and table:

Line	Mode	Del [s]	Vol [µl]	Flow [µl/s]	Time [s]	Duration [s]	Cycles	Feedback quantity	><	Value	DataN	Action	Pause [s]
1	F	1200	100.000	0.250	400.00	300		10A: O2 concentration [µM]	>	120.00	2	start	0
2	F	900	50.000	0.125	400.00	300		10A: O2 concentration [µM]	>	60.00	2	start	0
3	F	900	20.000	0.050	400.00	300		10A: O2 concentration [µM]	>	30.00	2	start	0
4	D	900	50.000	50.000	1.00	120	1						

1. Select a program line to edit by clicking on it. The selected program line in the program lines table is highlighted in blue. The current settings for this program line will be displayed in the input mask and can be edited. If you do not want to change the marked line but want to add a program line to the setup, click the control button **Insert** and a copy of the selected program line will be inserted in the program line table. If you want to add a program line at the end of the setup, click on **Append**.
2. Set the parameters for the program line using the input mask. A detailed description for the different modes is given below. The button **Restore program line** will appear in the input mask. The values in the input mask are reverted to those in the program line by pressing this button.
3. To transfer the new values from the input mask to the program line table either press **Replace** to overwrite the selected program line or **Insert** to insert a new program line before the selected program line with the values from the input mask. Only after pressing **Replace**, **Insert** or **Append** will the values in the program line table change. Selecting a different program line, without pressing **Replace** or **Insert** or **Append** will discard all changes!

The buttons **Replace**, **Insert** or **Append** will be highlighted in **yellow** when changes have been made in the input mask that have not yet been written to the program lines table.

The control buttons **Move up** and **Move down** can be used to change the position of a program line in the program line table and hence the sequence of the execution of program lines.


4.5. Parameters for a program line

This section describes the different parameters used in the program lines and how to use the input mask to set these parameters.

Common features: Some features are common in both direct and feedback control mode.

Delay [s] is the time [s] before starting the first instruction of any program line.

Injection volume, Flow, and Time: For each program line two of the three parameters injection volume, injection flow, and injection time have to be set. The third parameter is calculated from the two supplied by the user. Select one of the following three options for defining the volume, flow and time (duration) of active pumping:



Vol+Flow: Injection volume [µL] and flow [µL/s] are entered. If you change one value and click into another box, the corresponding time [s] for active pumping is calculated and displayed.

Vol+Time: Injection volume [µL] and time [s] are entered. If you change one value and click into another box, the corresponding flow [µL/s] is calculated and displayed.

Flow+Time: Injection flow [µL/s] and time [s] are entered. If you change one value and click into another box, the corresponding volume [µL] is calculated and displayed.

Error messages at zero volume/flow: If zero (0) is entered in either of the two editable fields of the 'TIP2k control' window, and you click on **Insert** or **Replace**, an error window appears.


Maximum volume and flow are shown depending on the syringe and TIP2k series used (200 µL or 500 µL syringe). If higher values are entered and you click on **Insert** or **Replace**, an error window appears (see below TIP2k configuration window).

☉ **TIP forward** is the standard operation mode for titrations and injections.

☉ **TIP backward** is selected if the pump should operate backwards. This is applied for automatically filling the TIP2k syringes. Be careful not to withdraw any sample from the O2k-chamber, by using the backward mode. This may clog the TIP2k needles. A warning is displayed before starting a TIP2k setup that includes a backward step.

Test start is used to activate a single operation of the TIP2k according to the volume/flow/time settings, without delay. The "Test start" mode does not interfere with the programmed TIP2k setup, i.e. it neither starts the first cycle nor reduces the total volume of the programmed titration-injection setup. After clicking on **Test start**, the button changes to **Test stop in yellow**, enabling to stop the pump before the injection is completed. It is advisable to use **Test start** before inserting the needles into the O2k-chamber. No events are set by DatLab when using Test start.

Mode selection: Some parameters differ between the modes for direct control and feedback control. Appropriate input masks are selected by pressing the **Direct control** or **Feedback control** buttons shown on the right.



4.6. Direct control mode:

In the direct control mode, a defined volume is injected into the O2k-chamber at a defined speed. This process can be repeated as desired in defined intervals. While one program line is executed, volume flow and interval stay constant. To achieve changes in these parameters (e.g. increasing volumes in later injection) several direct control program lines are run in sequence, one for each set of parameters.

Input mask: In the direct control mode only the upper left side of the control window is used for the input mask.

The screenshot shows the TIP2k control interface. The 'Direct control' mode is selected. The 'Volume [µL]' is set to 10.000, 'Flow [µL/s]' is 30.000, and 'Time [s]' is 0.33. The 'Interval [s]' is 180, 'Cycles' is 2, and 'Duration [s]' is 360. There are buttons for 'Test start' and 'Restore program line'.

Interval [s] is the total time from the start of a pumping cycle to the start of the consecutive cycle. It is the sum of the titration-injection (pumping) "Time [s]", and a pause between the end of the previous and the beginning of the following pumping state.

$$t_{\text{Interval}} = t_{\text{Pumping}} + t_{\text{Pause}}$$

In applications, the "Interval" is much longer than the pumping "Time", since the system has to attain a new steady state after the disturbance of the titration.

Cycles is the number of times the injection is repeated with a pumping phase followed by a pause.

Duration is calculated by DatLab from $\text{Duration} = \text{Interval} \times \text{Cycles}$. It is the total duration of the program line, excluding the delay.

After the entered values have been copied to the program lines table by Replace or Insert a direct control program line is described in the program line table by the parameters Del(ay), Vol(ume), Flow, Time, Duration, and Cycles. All other fields are not used in the direct control mode.

Line	Mode	Del [s]	Vol [µL]	Flow [µL/s]	Time [s]	Duration [s]	Cycles	Feedback quantity	><	Value	DataN	Action	Pause [s]
1	D	180	1.000	30.000	0.03	1080	6						
2	D	180	5.000	30.000	0.17	540	3						

In the picture above two program lines have been used to increase the volume per injection from 1 µL for the first six injections to 5 µL for the subsequent three injections. The expected final volume after all program lines is displayed below the program line table.

The screenshot shows a text box with the text "Predicted final volume: 21.000 µL". Below it are buttons for "Load setup", "Save setup", and "Hide details". A link "MitoPedia: Titration-Injection microPump" is also visible.

4.7. Feedback control mode

In the feedback control mode, the injection flow and - depending on usage - also time and volume for a single injection is set by the user. The total amount of volume injected into the O2k-chamber during the execution of the entire program line is, however, adapted by DatLab to maintain a certain level of a measured analytical parameter e.g.


oxygen concentration and / or to hold this parameter between predefined limits. Such a feedback control can be implemented in two ways:

Method A: Short injection pulses are programmed. Each injection is completely executed once started without regard to the measured analytical parameter. DatLab continues to start these injections as long as the start condition is met.


Method B: Injections are started when the start condition is met. If a stop condition is met while the injection is still running, the injection is stopped, even if the predefined volume has not been added yet. Therefore, only the flow speed is strictly maintained at the value set by the user but not time and volume.

Comparison: In method A, only one condition (the start condition) is used and the validity of the condition is checked only between injections but not during injections. In method B, two conditions are used (start and stop) and - while the validity of the start condition is also considered only between injections - the validity of the stop condition is checked during injections. By continuously monitoring and reacting to the analytical parameter the target can be reached more precisely (less overshoot) and faster (no breaks between single injections) than with method A. On the other hand, the pauses between injections following method A allow the system to reach equilibrium before the decision for or against a further injection.


Input mask: The input mask for feedback control utilizes additionally the upper right part of the control window for feedback control-specific parameters.



Injection Volume, Flow and Time are entered as described above. However, if a stop criterion is used only the **Flow** will be maintained absolutely. **Time** and **Volume** are rather maximal values for each individual injection, i.e. an injection is stopped when either the maximum volume has been reached or a stop criterion is achieved. These are limits valid for each individual injection. During executing one program line, many injections may be started, and the total volume added may be higher. In the feedback mode, it is not possible to predict what the total volume will be, therefore below the program line table the message "Feedback control - cannot predict final volume" is printed.



Duration of feedback control: Above the table containing feedback conditions are two fields for determining how long the program line should be executed, i.e. how long the feedback control should be maintained.



To maintain feedback control for x seconds only fill in the first field. The second field does not set an absolute time limit but stops the program line after y injections have been executed. The latter option will probably be useful only in special applications. Setting a parameter to zero removes any limit; therefore, setting both parameters to zero will maintain feedback control indefinitely.

Feedback conditions: The heart of feedback control is setting conditions that will start or terminate an injection. This is done in the **feedback conditions table**. The following paragraphs describe the fields used for defining feedback control. How

these parameters may be used to implement feedback control is described in **Implementing Feedback Control** below.

There will always be at least one feedback condition line visible, more may be generated by pressing **Insert** or removed by pressing **Delete**.

Some fields are selection fields – a drop-down button ▼ will appear when clicking on the field.

Quantity	><	Value	DataN	Action
1A: O2 concentration [µM]	>	120	2	Start
1A: O2 concentration [µM]	<	100	1	Stop

Quantity	><	Value	DataN	Action
10A: O2 concentration [µM]	<	10	1	Start
10A: O2 concentration [mM]				
10A: O2 concentration [nM]				
10A: O2 pressure kPa [kPa]				
10A: O2 pressure mmHg [mmHg]				
10A: O2 slope neg. [mV/s]				
10A: O2 slope neg. [pmol/(s*ml)]				
10A: O2 flux per V [pmol/(s*ml)]				

Quantity is a selection field. Select the analytical parameter that you want to control from the drop-down menu.

>< is a selection field. It determines whether the condition is met when the measured analytical parameter is higher or lower than **Value**. Select > or < from the drop-down menu.

Value is the numerical value for the chosen analytical parameter you want to use as condition. The unit is given in the field **Quantity**. Enter a numerical value here. The fields **Quantity**, **><**, and **Value** together define a **condition**.

DataN is the number of data points for which the **condition** must be met before **Action** is executed. Entering higher numbers (>1) here allows the system to stabilize, preventing injections being triggered by non-significant fluctuations and allowing the system to equilibrate after an injection.

Action is a selection field. Select "Start" or "Stop" from the drop-down menu. "Start" will start an injection with the parameters entered in the left side of the input mask when the **condition** is met for **DataN** points. When an injection is running, the program will continually check the feedback conditions table for a line containing a "Stop". If the **condition** in such a "Stop" line is met for **DataN** data point the injection will be terminated. Method A uses only "Start" actions, method B uses "Start" and "Stop" actions.

Pause [s] is used to set a minimal pause between injections. No injection will be started for this time after an injection was finished, even if the start condition is met. This feature is only meaningful for method A.

Implement feedback control

Method A Only "Start" actions are used. A feedback condition line is constructed defining only one (upper or lower) limit for the analytical parameter.

Injection **Volume**, **Flow**, and **Time** are set to achieve rather short pulses.

After the injection, the **condition** is checked again. When the limit is

still not reached (the **condition** is still fulfilled) after an optional delay of **Pause** seconds, a second injection identical to the first is performed. This is repeated until the limit is reached (the **condition** is no longer fulfilled). Then the pump is inactive until the **condition** is fulfilled again. How the **condition** should be defined in respect to the limit (> or <) depends on the nature of the process in the O2k-chamber and of the agent used in the syringes to counteract this process. If for example the analytical parameter used for feedback control is oxygen concentration, the concentration will usually decrease due to respiratory activity. This can be counteracted by filling the TIP2k syringes either with oxygen saturated media or with hydrogen peroxide (when catalase is present in the medium). The start **condition** will therefore define a lower limit for oxygen concentration and the operator < will be used. So, the line can be expressed as:

If $c(\text{O}_2) < \text{lower limit}$, then start injection.

Using this method there is no programmed upper limit to oxygen concentration but a *de facto* upper limit will be reached by setting the injection time and volume to appropriate values.

Method B "Start" and "Stop" Actions are used. One feedback condition line is constructed defining the first limit and a "Start" Action. A second feedback condition line is constructed defining a second limit and a "Stop" Action.

The injection **Flow** is set to a rather low value, allowing on-line monitoring of the analytical parameter during the injection.

Injection **Volume** or **Time** have to be set to some maximal value but usually will have no consequences in this method.

When the condition defined in the line containing the "Start" Action is fulfilled, an injection will start. This injection will be stopped when the second

Quantity	><	Value	DataN	Action
* O2 Concentration [W] [nmol/ml]	<	10	2	Start

Vol+Flow	Volume [µl]	5.000	<input type="radio"/> TIP backward
Vol+Time	Flow [µl/s]	30.000	<input checked="" type="radio"/> TIP forward
Flow+Time	Time [s]	0.17	Test start

Quantity	><	Value	DataN	Action
1A: O2 concentration [µM]	>	120	2	Start
1A: O2 concentration [µM]	<	100	1	Stop

Vol+Flow	Volume [µl]	5.000	<input type="radio"/> TIP backward
Vol+Time	Flow [µl/s]	0.010	<input checked="" type="radio"/> TIP forward
Flow+Time	Time [s]	500.00	Test start

limit is reached as defined in the line containing the **"Stop" Action**. The pump will then be inactive until the condition for the **"Start" Action** is fulfilled again. Which of the two limits should to be placed in the start and stop lines depends again on the nature of the process in the O2k-chamber and the content of the TIP2k syringes. In the example given for method A (oxygen concentration being controlled, respiratory activity in the O2k-chamber, and e.g. peroxide in the syringes) the line containing a **"Start" Action** will set the lower limit for oxygen concentration, while the line containing the **"Stop" Action** will define the higher limit for oxygen concentration. Thus, oxygen concentration will be held between an upper and lower limit. This may be formulated as:

If $c(O_2) < \text{lower limit}$, then start injection.


If $c(O_2) > \text{upper limit}$, then stop injection.

4.8. Save, load and manage TIP2k setups

Save: After modifying the TIP2k control settings, the new configuration may be saved. To indicate this possibility the button, **Save setup** will be highlighted in **yellow** after changes were done to the current setup. To apply a new name to the setup, type the new name into the setup name field on the left side of the save button and then press **Save setup**.




Load setup: A new setup is selected by activating the drop-down menu in the setup name field (click on ▼). After a new setup was selected the




Load setup button will be highlighted in **yellow**. Clicking on **Load setup** will then load the new setup into the 'TIP2k control' window. Any unsaved changes to the current setup will be lost. The **Load setup** and **Save setup** buttons are available from all tabs of the 'TIP2k control' window.

Manage setups: To delete or rename setups use the Manage setups function. This is accessible from the main menu under **TIP2k**. A list with all currently available setups will appear. These may be renamed or deleted.



Import Oroboros setups: Additional TIP2k setups supplied by Oroboros Instruments in a template archive can be imported by choosing **Import DatLab templates** from the **File** Menu ([MiPNet19.01C](#)).



Transferring setups: Setup files (name.TCS) are stored in the directory /DatLab/APPDATA/TIPCTRL. Any "name.TCS" file manually copied into this directory will be recognized as a TIP2k setup file. This may be used e.g. to manually transfer a TIP2k setup written on one computer to another computer.

4.9. Running a TIP2k setup program

Start: By clicking on Start, the TIP2k setup program is started as displayed in the 'TIP2k control' window and in the status line.



This does not necessitate that an actual injection is started at the same time. The program line being executed will be highlighted in blue in the program line table. When an injection is started, this is displayed in the status line, additionally a green light on the button on the TIP2k unit blinks.

Stop immediately: At any time, the entire TIP2k program may be terminated by pressing **Stop immediately**. If an injection is running, this injection will be stopped. The total volume injected is given in a TIP2k event, see below. If the TIP2k setup is started again afterwards, it will be started at the first program line! Therefore, if the program should be continued at a certain point after a stop instruction, the setup has to be changed accordingly.

Suspend will halt the progress of the current program line if no injection is taking place at the moment. The time counter in the status line will show elapsed time for the current program line, but no injection will be started. A running injection will **NOT** be stopped by Suspend.

Resume will continue a setup halted with **Suspend**. If the scheduled time for an



injection was reached during the time the TIP2k suspended this injection will start immediately after pressing **Resume**.

The buttons **Repeat** and **Next**, described below, are only available while the program is in the **Suspend** mode. They allow modifying a running TIP2k program without having to stop and restart it from line 1.

Repeat will repeat the last performed injection and resume the TIP2k program. Example: The TIP2k was programmed for three 1 μ L injections and then one 10 μ L injection. By suspending the program after the third injection and clicking **Repeat**, a fourth 1 μ L injection is done before the program continues with the 10 μ L injection.

Next will immediately jump to the next action (within the line, or to a new line from the end of the line). This may be an injection, a delay or a feedback control phase. Example: Several injections with intervals of 15 minutes were programmed. Five minutes after the second injection, the observed signals are already stable. By suspending the program and then clicking **Next** the TIP2k immediately executes the third injection.

Close will close the control window. When pressed while a TIP2k set up is executed, the setup will not be stopped, but will run in the background. By selecting 'TIP2k control' from the TIP2k menu again [F8] the TIP2k window will be opened again, displaying the progress of the TIP2k setup and giving access to the control buttons **Stop immediately** and **Suspend**. Therefore, the 'TIP2k control' window can be closed after starting the setup and other DatLab features can be used with the setup running in the background.

4.10. Chemicals window

The Chemicals window is selected by pressing on the Chemicals tab always visible in the TIP2k window.


Left/Right refers to the side of the syringes and O2k-chambers.

Solvent is the liquid in which the chemical in the TIP2k syringe is dissolved.

Substance is the chemical in the TIP2k syringe to be titrated into the O2k-chamber.



Conc. in TIP refers to the stock concentration [mM or μM] of the chemical dissolved in the TIP2k syringe.

The information written into the Info window is saved with the TIP2k setup.




4.11. TIP2k configuration window

The TIP2k configuration window is always accessible by pressing on the 'Configuration' tab in the TIP2k window. It displays the TIP2k syringe volume [μL] and the TIP2k series. DatLab 7 will give an alert if you try to titrate more volume than possible. This window is for information only.



4.12. Info window

By clicking on the 'Info' tab, the Info window is displayed. Here a description of the TIP2k setup can be entered. There is also a field for an URL address that may contain a web link. This will be used to supply additional information for TIP2k setups provided by Oroboros



Instruments. The information written into the 'Info' window is saved with the TIP2k setup and can be viewed after loading a TIP2k setup.

4.13. TIP2k status


The current status of the TIP2k is indicated in the O2k status line at the bottom of the screen.



TIP: Ready appears in the status bar, when the TIP2k is properly connected to the O2k, the Oroboros O2k is switched on and connected to the PC with the RS232 cable, and DatLab is connected [F7].

TIP: Test start: the TIP performs a pump operation after clicking on Test start, without setting an event. The background color is green if the TIP2k pumps in the "TIP forward" mode, and red if it pumps in the "TIP backward" mode.

Setup running: When a setup is running but no injection is taking place, the status line will be highlighted in yellow.



Information being displayed includes:

Tip Line: the program line being executed, e.g. **TIP Line 1**

mode: the modus used in the current program line, e.g. **Direct**

Vtot: total volume injected until the end of the last executed injection (volume will be updated after the end of each injection), e.g. **Vtot: 1.000 µL**

Cyc: number of cycles executed / total number of cycles, e.g. **Cyc. 2/3** (total number of cycles is not known in feedback control)

Time: time (of current activity) evolved / total time of current activity, e.g. **67/100 s** (total time is not known in feedback control)

Pump running: While an injection is executed, similar information as described above are displayed, but the TIP2k status is highlighted in green.



TIP suspended: When the progress of a TIP2k setup is suspended the available information are displayed, the time counter continues to display the evolved time followed by the word "suspended". The status line is highlighted in yellow.



Special operations:

TIP: Manual forward - the pump is moved forward manually, using the front push button of the TIP2k. This status is highlighted with a green background.

TIP: Manual back - the pump is moved backward manually, using the rear push button of the TIP2k. This status is highlighted with a red background.

TIP Errors: are highlighted in red, additionally a red error field is displayed in the status line. The error message can be cleared by clicking on this error field.



TIP: Rear limit - appears with a red background when the rear limit is reached upon moving the pump backward.

TIP: Front limit - appears with a red background when the front limit is reached upon moving the pump forward.

4.14. TIP2k events

After **Start**, "Events" are set automatically in the graph as vertical lines with an "Event name" at the top of the graph. Note no events are set with **Test start**.

Click on the event name, at the top of the graph, to view the event text which automatically reports additional information.

TIP event names

TIP start is an event marking the start of a new TIP2k setup program.

P001: marks the beginning of the first actual pumping event (the first injection). The programmed volume and the injection speed will be displayed in the event text. In direct control mode and in feedback control mode without using stop conditions the displayed volume will correspond usually to the injected volume (unless the injection was stopped manually). In feedback mode with using stop condition this initially programmed volume usually will not be reached. The correct volume in this case (and in the case of a manual termination) is displayed in the **PStop** event. Injections are numbered sequentially **P002, P003**,... during the entire TIP setup program. If the TIP2k is operated in the "TIP backward" mode, the events set by the TIP2k are the same as in the pumping (TIP forward) mode, but the volume indicated in the event text has a negative sign.


PStop: is an event set at the end of an injection, if

- the injection was stopped manually by the user,
- the injection was stopped by a feedback control stop condition,
- the entire programmed volume was injected, and injection time was ≥ 30 s.


This event indicates the time when the actual pumping has switched off. In the event text the volume actually injected is displayed.

Summary: injected volumes: When a **PStop** event is displayed, the corresponding event text will always contain the correct (actually injected) volume. When no **PStop** event is displayed the **P00x** event will contain the correct (actually injected) volume. There will be no **PStop** event for injections that were carried out until the entire programmed volume was injected and lasted less or equal to 30 seconds.

TIP end: appears at the end of the TIP2k setup program (after the last program line has been executed or the setup program was stopped by the user). In the event text the final volume injected during the entire setup program is displayed.



TIP Error: appears when the programmed injection is stopped due to mechanical limits of the TIP2k, when the rear or front limits are reached. Simultaneous with this event the status bar of the TIP2k changes to a red background with the text “TIP: Front limit” or “TIP: Rear limit”. Click on the Error event name, to view the text with information on the volume actually pumped.



A TIP2k error will automatically stop the current program (event: **TIP end**). An additional event ERROR without the volume information may be generated, depending on software version.

In addition to the TIP Error event in the graph, the green control light on the push button of the TIP2k blinks rapidly when “TIP: Front limit” or “TIP: Rear limit” is reached. At this status you cannot operate the TIP2k with DatLab. In order to activate the TIP2k for further use, the pump has to be reset manually, using the push buttons. Then the status “TIP: Ready” is attained, the green light stops blinking and is continuously on.

5. References

Gnaiger E (2001) Bioenergetics at low oxygen: dependence of respiration and phosphorylation on oxygen and ADP supply. *Respir Physiol* 128: 277-97. »[Bioblast link](#)



Further reading:

O2k-Manual

» [MiPNet26.06 DatLab 7: Guide](#)

6. Author contributions

Mario Fasching contributed to this MiPNet as a former member of Oroboros Instruments.

7. Acknowledgements

Supported by the NextGen-O2k project.



This project has received funding from the European Union's Horizon 2020 research and innovation programme under grant agreement No 859770.

